


Shape of groove's cross-section not important, but round shape is preferred. Also, the grooves must be at least 0.250mm wide and deep.



Groove may continue through entire length of needle.

PROPRIETARY AND CONFIDENTIAL

THE INFORMATION CONTAINED IN THIS DRAWING IS THE SOLE PROPERTY OF STANFORD UNIVERSITY. ANY REPRODUCTION IN PART OR AS A WHOLE WITHOUT THE WRITTEN PERMISSION IS PROHIBITED.

UNLESS OTHERWISE SPECIFIED:

DIMENSIONS ARE IN MM
TOLERANCES: 0.01

INTERPRET GEOMETRIC
TOLERANCING PER:

MATERIAL
Titanium Alloy

FINISH

DO NOT SCALE DRAWING

NAME

DATE

DRAWN Santhi Elayaperumal 7/25/08

CHECKED

ENG APPR.

MFG APPR.

Q.A.

COMMENTS:

TITLE:

16ga Stylet

DWG. NO.

16gaStylet

REV

SIZE
A

WEIGHT:

SHEET 1 OF 1